



ICAR- KRISHI VIGYAN KENDRA
(ICAR-INDIAN INSTITUTE OF HORTICULTURAL RESEARCH)



GONIKOPPAL-571213, Kodagu, Karnataka

Phone/Fax: 08274-247274

E-mail:ihrkvkgk@gmail.com

F. No: KVK/2017-18/

Date: 01.01.2018

Quotations are invited for supply of Vertical Autoclave as per the following

Name of the equipment	SPECIFICATIONS	Quantity
Vertical Autoclave	<ul style="list-style-type: none">• Inner chamber made out of 2 mm thick stainless steel 304 sheet, Outer chamber is made up of 1 mm thick stainless steel 304 sheet, Lid is made out of 6 mm thick stainless steel 304 plate, Flange is made out of 6 mm thick stainless steel 304 Flat, Fitted with one pressure gauge, one safety valve, Working pressure of the autoclave is 10 to 30 psi.• The autoclave is tested hydraulically upto 40 psi.• Micro processor based PID Digital temp. controller with inbuilt timer• Inner chamber dimension : (Dia x Depth) 24" x 36"• Power rating : 6 Kw Cap.260 ltrs.• Perforated stainless steel basket : 3 No.	1 No.

1. Quotation should be inclusive of all packing, forwarding, insurance, freight etc. In case Quotation is exclusive of the above amount on this account should be indicated separately.
2. Complete descriptions, specifications and make of the items/accessories should be furnished along with the Quotation. Necessary literature and pamphlet if any should also be sent along with the Quotation. Rate per unit should be given precisely and delivery period clearly mentioned.
3. The rates quoted should be FOR destination and should be kept open for 180 days from the date of Quotation. Installation/handing over is required to be executed at Krishi Vigyan Kendra, Gonikoppal, Kodagu District - 571 213.
4. Warranty period, conditions of service after sale, annual maintenance charges, supply of spare parts, if any should also be clearly mentioned.
5. Payment will be arranged only after supply, satisfactory installation of the equipment and approval.
6. The Quotation addressed to the **Senior Scientist & Head, Krishi Vigyan Kendra, Gonikoppal, Virajpet Taluk, Kodagu District - 571 213, Karnataka** state should reach on or before **20-01-2018**, by post in a sealed cover subscribing **Quotation for the supply of Vertical Autoclave**.
7. The right to accept or reject in full or part of the Quotation is reserved with Head of station.

Yours faithfully,

Sd/-

SENIOR SCIENTIST & HEAD