

Corrigendum with respect to GCMS/MS Specifications

Sl. No	Requirement	Specification
5.	Triple Quadruple Mass Spectrometer	EI MRM S/N: 15,000:1 for 100fg of transition of m/z 272 & 222 using 30m x 0.25mm ID x 0.25um column <u>MAY BE READ AS</u> EI MRM S/N: 10,000:1 or more for 100fg of transition of m/z 272 & 222 using 30m x 0.25mm ID x 0.25um column